

**Angio Upper Extremity**

Siemens Flash

Application Examples: cold finger

Oral Contrast	No
IV Contrast / Volume / Injection Rate	Omnipaque 350 / 100 mL / 5 mL/sec
MedRad Protocol	Angio Upper Extremity

*Technical Factors*

Care Bolus ROI Location / HU	Aortic Arch / 130
Monitoring Delay	10 seconds
Cycle Time	1.43 seconds
Scan Delay	3 seconds
Patient Instructions	Do not swallow

Scan Type	Spiral
Detector Collimator	Acq 128 x 0.6 mm
Care kV	On / 120 kV
Care Dose 4D	On / 160mAs
Rotation Time (seconds)	0.28
Pitch	1.2
Typical CTDIvol	10.82 mGy $\pm$ 50%

Topogram: Lateral and AP, 768 mm

CTA Upper Extremity	Recon Type	Width / Inc	Algorithm	Safire	Window	Series Description	Networking	Post Processing
Recon 1	Axial	1 x 1	I30f	2	Angio	AXIAL	PACS	None
Recon 2	3D:COR	1 x 1	I30f	2	Angio	COR MIP	PACS	Coronal MIP
Recon 3	3D:SAG	1 x 1	I30f	2	Angio	SAG MIP	PACS	Sagittal MIP
Recon 4	Axial	0.6 x 0.6	I26f	2	Angio	AXIAL 0.6 STND	TR & PACS	None

**IV Placement:** *preferably* 18gauge in antecubital (AC) fossa of unaffected arm.**Patient Preparation:** Have patient remove any detachable dental work.**Patient Position:** Patient lying supine with unaffected arm at side. Tuck chin slightly and position head so the sella is parallel to the gantry in a symmetrical position (no rotation or tilt) with neck in neutral position. Put affected arm above head as straight as possible.**Scan Range:** Include aortic arch through fingers.**Scan Instructions:** Place pre-monitoring ROI in aortic arch.**Recons and Reformations:** Make coronal and sagittal 1x1 MIPs.