

## ***Rapid Stroke Head (TPA)***

Application Examples: acute stroke

Sequence	FoV	Thickness	Gap
AX DWI (B-0, B-1000)	24CM	3MM	1MM
AX T2 FLAIR PROPELLER	24CM	4MM	1MM
AX MERGE	24CM	4MM	1MM

**GET PATIENT OFF TABLE AND CONTACT NEURORADIOLOGIST IMMEDIATELY!**



AXIAL Coverage